


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2038, Rack A, Plate 1, Well C02, 489993
[CC778DF652EA26C2] Jan 09, 2026 11:27:22 Pressure = 81 bar (80 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.49	1.0%	10373	
2	0.497	F 0.59	0.6%	5963	
3	0.703	F 0.84	15.0%	161384	Acetylated F peak
4	0.843	F 1.01	67.4%	722453	Consistent with F
5	1.205	A 1.01	16.1%	172286	A peak

Total Area: 1072459

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range