


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2110, Rack A, Plate 1, Well B10, 498865
[9D27402FB189F995] Mar 31, 2026 11:44:11 Pressure = 59 bar (59 to 61)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.48	1.8%	7850	
2	0.490	F 0.56	2.4%	10443	
3	0.723	F 0.83	0.6%	2487	
4	0.877	F 1.01	13.3%	57746	Consistent with F
5	1.052	F 1.21	6.5%	28335	3
6	1.253	A 1.01	73.6%	320040	A peak - REVIEW
7	1.425	S 0.90	1.8%	7657	3
Total Area:				434558	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range