


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well G06, 493131
[CD38D2DFB77DB11D] Jul 15, 2026 14:56:33

Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.692	F	0.83	20.2%	350014	Acetylated F peak
2	0.842	F	1.01	70.8%	1224677	Consistent with F
3	1.167	A	1.01	9.0%	156007	A peak
Total Area: 1730698						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range