


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2041, Rack A, Plate 1, Well F02, 171862
[8D27882306C88CE8] Jan 15, 2026 12:53:33 Pressure = 81 bar (81 to 84)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.49	0.4%	8104	
2	0.508	F 0.60	0.3%	4744	
3	0.695	F 0.82	13.5%	244355	Acetylated F peak
4	0.825	F 0.97	75.0%	1361141	Consistent with F
5	1.207	A 1.02	10.9%	197643	A peak
Total Area:				1815987	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range