


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2025, Rack A, Plate 1, Well A01, 490104
[C862C47DC4FD5115] Dec 23, 2025 11:02:59 Pressure = 78 bar (77 to 80)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.48	0.8%	6901	
2	0.502	F 0.59	0.5%	4101	
3	0.708	F 0.83	15.4%	138952	Acetylated F peak
4	0.845	F 1.00	74.8%	673641	Consistent with F
5	1.232	A 1.02	8.6%	77200	A peak
Total Area:				900795	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range