


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2169, Rack A, Plate 1, Well A09, 501545
[8D3CA8BE4D3846C0] Jun 05, 2026 13:27:55 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.422	F	0.50	0.7%	7956	
2	0.507	F	0.60	0.8%	8480	
3	0.717	F	0.85	14.2%	152021	Acetylated F peak
4	0.833	F	0.99	59.4%	636897	Consistent with F
5	1.200	A	1.01	24.7%	265081	A peak - REVIEW
6	1.372	S	0.88	0.2%	2496	
Total Area: 1072931						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range