


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2169, Rack A, Plate 1, Well F03, 493851  
[D86CBCFE47B0C442] Jun 05, 2026 15:21:06 Pressure = 64 bar (64 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.49	0.7%	10376	
2	0.512	F 0.59	0.6%	8211	
3	0.705	F 0.82	12.9%	183398	Acetylated F peak
4	0.852	F 0.99	60.4%	858308	Consistent with F
5	1.197	A 1.01	25.4%	361898	A peak - REVIEW
Total Area: 1422191					

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range