


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2204, Rack A, Plate 1, Well F10, 487183
[9D3989E96D1F93BF] Jul 23, 2026 12:34:36 Pressure = 93 bar (93 to 95)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.522	F	0.62	0.8%	13065	
2	0.707	F	0.83	13.3%	230930	Acetylated F peak
3	0.815	F	0.96	75.9%	1321469	Consistent with F
4	1.163	A	1.01	10.1%	175901	A peak
Total Area: 1741365						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range