


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2107, Rack A, Plate 1, Well G09, 498225
[996241D4005E0C62] Mar 25, 2026 13:57:26 Pressure = 57 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.47	1.2%	13062	
2	0.510	F 0.58	0.8%	8218	
3	0.730	F 0.83	14.7%	154121	Acetylated F peak
4	0.877	F 0.99	66.2%	694951	Consistent with F
5	1.263	A 1.01	17.0%	178843	A peak
Total Area:				1049195	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range