


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1919, Rack A, Plate 1, Well G11, 170327
[8D2042C155D6F1BF] Sep 24, 2025 13:39:56 Pressure = 60 bar (59 to 61)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.412 | F 0.49 | 0.7% | 9401 | |
| 2 | 0.495 | F 0.58 | 0.8% | 9678 | |
| 3 | 0.682 | F 0.80 | 13.9% | 176143 | Acetylated F peak |
| 4 | 0.832 | F 0.98 | 68.0% | 861482 | Consistent with F |
| 5 | 1.207 | A 1.01 | 16.6% | 209812 | A peak |
| Total Area: | | | | 1266516 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range