


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2149, Rack A, Plate 1, Well C05, 501884  
[DD78A007A125B9BF] May 20, 2026 13:05:16 Pressure = 65 bar (65 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.60	1.8%	21149	
2	0.717	F 0.84	12.6%	151469	Acetylated F peak
3	0.865	F 1.01	58.4%	700156	Consistent with F
4	1.098	A 0.91	0.2%	2137	
5	1.212	A 1.00	27.0%	324370	A peak - REVIEW
Total Area:				1199281	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range