


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2133, Rack A, Plate 1, Well D08, 502919
[98330AEC3A1CA4E2] Apr 29, 2026 12:38:06 Pressure = 65 bar (64 to 66)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.512	F	0.59	3.2%	49225	
2	0.717	F	0.83	18.8%	292463	Acetylated F peak
3	0.863	F	1.00	65.1%	1013972	Consistent with F
4	1.220	A	0.99	13.0%	203022	A peak
Total Area: 1558682						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range