


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well G03, 487164
[8C28878AE22ACE68] Jul 15, 2026 14:50:29 Pressure = 84 bar (83 to 86)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.692	F	0.83	19.3%	414378	Acetylated F peak
2	0.812	F	0.97	72.2%	1553255	Consistent with F
3	1.162	A	1.01	8.6%	184048	A peak
Total Area: 2151681						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range