


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well A09, 502241
[9D68EC72D3926E62] Jun 10, 2026 13:17:42 Pressure = 67 bar (67 to 69)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.61	2.2%	23238	
2	0.710	F 0.85	16.9%	179198	Acetylated F peak
3	0.830	F 0.99	67.8%	718540	Consistent with F
4	1.203	A 1.01	13.1%	138624	A peak
Total Area: 1059600					

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range