


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2003, Rack A, Plate 1, Well D12, 483942  
 [D972BBF1729AA4CA] Dec 03, 2025 11:44:40 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.9%	10831	
2	0.498	F 0.58	0.6%	6914	
3	0.727	F 0.84	16.9%	202710	Acetylated F peak
4	0.843	F 0.98	69.6%	836052	Consistent with F
5	1.012	F 1.17	0.9%	10507	
6	1.237	A 1.02	11.2%	134798	A peak
Total Area:				1201812	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range