


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2080, Rack A, Plate 1, Well F08, 505732
 [CC363352F6040468] Feb 24, 2026 12:43:17 Pressure = 56 bar (56 to 58)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	2.5%	14367	
2	0.508	F 0.58	1.2%	6892	
3	0.733	F 0.83	3.8%	21622	
4	0.783	F 0.89	1.4%	7911	
5	0.872	F 0.99	80.2%	454352	Consistent with F
6	1.265	A 1.01	10.9%	61540	A peak
Total Area:				566684	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range