


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well H06, 493012
[D979850AE320CE68] Jul 09, 2026 13:25:09

Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.690	F	0.83	19.3%	334223	Acetylated F peak
2	0.818	F	0.99	73.6%	1273795	Consistent with F
3	1.150	A	1.00	7.1%	122097	A peak
Total Area: 1730115						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range