


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2092, Rack A, Plate 1, Well A05, 498801  
[D9772FE2D899591F] Mar 11, 2026 11:32:16 Pressure = 58 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.418	F 0.48	9.8%	86595	
2	0.728	F 0.83	16.1%	142019	Acetylated F peak
3	0.880	F 1.01	58.4%	513705	Consistent with F
4	1.245	A 1.00	15.7%	138048	A peak

Total Area: 880367

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range