


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2049, Rack A, Plate 1, Well E09, 490380
[8D77894FBE482EC8] Jan 20, 2026 12:54:17 Pressure = 80 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.6%	10834	
2	0.507	F 0.60	0.2%	1580	
3	0.703	F 0.83	13.9%	93271	Acetylated F peak
4	0.845	F 1.00	77.8%	523998	Consistent with F
5	1.208	A 1.01	6.5%	43471	A peak
Total Area:				673154	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range