


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well F12, 486003
[CD3CAC62FB984C4A] Jun 10, 2026 15:29:04 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.423	F	0.50	1.0%	8946	
2	0.502	F	0.59	0.3%	3180	
3	0.753	F	0.89	11.9%	110973	Acetylated F peak
4	0.840	F	0.99	65.1%	609343	Consistent with F
5	1.220	A	1.03	21.7%	203049	A peak - REVIEW
Total Area: 935491						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range