


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2197, Rack A, Plate 1, Well A09, 487107
[9D28879FBF7F3BB5] Jul 15, 2026 12:33:03 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.513	F 0.62	1.5%	8949	
2	0.680	F 0.82	12.7%	76770	Acetylated F peak
3	0.838	F 1.01	61.3%	370522	Consistent with F
4	1.165	A 1.01	24.5%	147960	A peak - REVIEW
Total Area: 604201					

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range