


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2061, Rack A, Plate 1, Well H04, 483350
 [9973CB05F41DF135] Jan 30, 2026 13:01:27 Pressure = 88 bar (88 to 90)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.48	0.9%	11395	
2	0.503	F 0.60	0.5%	6103	
3	0.702	F 0.83	9.3%	117091	Acetylated F peak
4	0.837	F 0.99	76.9%	966523	Consistent with F
5	1.208	A 1.01	12.4%	155624	A peak
Total Area:				1256736	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range