


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2025, Rack A, Plate 1, Well E06, 490116  
[D876906DEC77FB35] Dec 23, 2025 12:52:07 Pressure = 75 bar (75 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.48	1.2%	10132	
2	0.498	F 0.59	1.1%	9634	
3	0.713	F 0.84	12.2%	106472	Acetylated F peak
4	0.860	F 1.01	53.1%	462791	Consistent with F
5	1.000	F 1.17	2.5%	21709	
6	1.222	A 1.00	29.9%	260812	A peak - REVIEW
Total Area:				871550	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range