


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1957, Rack A, Plate 1, Well C09, 170558
[98641E7F99E9A4CA] Oct 24, 2025 10:40:40 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.390	F 0.46	1.3%	9270	
2	0.505	F 0.59	0.4%	2724	
3	0.705	F 0.82	15.5%	106881	Acetylated F peak
4	0.845	F 0.99	61.5%	424100	Consistent with F
5	0.995	F 1.16	0.5%	3534	
6	1.232	A 1.00	20.7%	142756	A peak - REVIEW
Total Area:				689265	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range