


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2038, Rack A, Plate 1, Well E02, 488958
[992788F650CA2E42] Jan 09, 2026 12:17:54 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.502	F 0.59	1.0%	25627	
2	0.702	F 0.83	16.6%	431506	Acetylated F peak
3	0.830	F 0.98	71.0%	1845141	Consistent with F
4	1.203	A 1.00	11.4%	296812	A peak
Total Area:				2599086	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range