


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1850, Rack A, Plate 1, Well B12, 168881  
[8871CDD328C0599D] Jul 02, 2025 11:15:21 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.398	F 0.47	1.9%	17092	
2	0.488	F 0.58	2.8%	25005	
3	0.713	F 0.85	13.0%	116228	Acetylated F peak
4	0.858	F 1.02	55.6%	496661	Consistent with F
5	1.102	A 0.91	0.3%	2646	
6	1.218	A 1.01	26.4%	235358	A peak - REVIEW
Total Area:				892990	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range