


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2182, Rack A, Plate 1, Well D09, 501423
 [CC6CAA2DC432E640] Jun 23, 2026 11:40:03 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.61	0.8%	20361	
2	0.697	F 0.82	13.9%	352109	Acetylated F peak
3	0.818	F 0.96	75.9%	1917409	Consistent with F
4	1.177	A 1.00	9.3%	236023	A peak
				Total Area: 2525902	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range