


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2188, Rack A, Plate 1, Well F03, 486275
[9C78D0967D224C4A] Jul 01, 2026 13:01:52 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.520	F 0.62	0.9%	11652	
2	0.692	F 0.83	13.9%	180311	Acetylated F peak
3	0.818	F 0.98	67.8%	877931	Consistent with F
4	1.163	A 1.00	17.3%	224476	A peak
Total Area:				1294370	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range