


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2065, Rack A, Plate 1, Well H10, 492212
[CC6624AAC704244A] Feb 04, 2026 14:52:16 Pressure = 84 bar (84 to 86)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.50	0.9%	9483	
2	0.510	F 0.61	0.2%	1699	
3	0.692	F 0.83	11.6%	118809	Acetylated F peak
4	0.835	F 1.00	73.2%	751898	Consistent with F
5	1.177	A 1.00	14.2%	145721	A peak
Total Area:				1027610	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range