


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1974, Rack A, Plate 1, Well E07, 170994
[DC67F3BB308DFB9F] Nov 11, 2025 13:00:40 Pressure = 66 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.390	F 0.45	2.6%	19066	
2	0.493	F 0.57	1.0%	7549	
3	0.728	F 0.84	19.8%	144956	
4	0.880	F 1.02	61.6%	452146	Consistent with F
5	1.237	A 1.00	15.0%	110127	A peak
Total Area:				733844	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range