


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2046, Rack A, Plate 1, Well A12, 171843
 [DD76886C399F5317] Jan 16, 2026 10:24:18 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.48	1.0%	9324	
2	0.505	F 0.60	0.4%	3524	
3	0.690	F 0.82	14.4%	129761	
4	0.832	F 0.99	67.1%	605014	Acetylated F peak Consistent with F
5	0.985	F 1.17	0.9%	7683	
6	1.207	A 1.02	16.2%	146034	A peak
Total Area:				901340	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range