


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well E01, 493021
[DC2D92389977933F] Jul 17, 2026 15:55:00

Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.692	F	0.83	20.8%	379816	
2	0.815	F	0.97	66.9%	1218761	Consistent with F
3	1.150	A	1.00	12.3%	224437	A peak
Total Area: 1823014						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range