


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1953, Rack A, Plate 1, Well A02, 859807  
[C8700A9F9741A4E0] Oct 21, 2025 10:17:19 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.395	F 0.46	1.2%	7817	
2	0.505	F 0.59	0.4%	2735	
3	0.717	F 0.84	12.9%	86502	Acetylated F peak
4	0.857	F 1.00	58.8%	394022	Consistent with F
5	0.993	F 1.16	1.5%	10111	
6	1.228	A 1.00	25.3%	169473	A peak - REVIEW
Total Area:				670660	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range