


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1359, Rack A, Plate 1, Well D10, 27485  
[DCCDA51274F262CA] Apr 03, 2024 11:06:00 Pressure = 61 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.487	F 0.55	9.2%	72428	
2	0.747	F 0.85	2.9%	22781	
3	0.787	F 0.89	0.6%	4594	
4	0.888	F 1.01	66.4%	520732	Consistent with F
5	1.252	A 1.00	20.9%	163963	A peak - REVIEW
Total Area:				784498	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range