


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2146, Rack A, Plate 1, Well A02, 496436
 [997DA18082D246EA] May 13, 2026 11:43:24 Pressure = 67 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.525	F 0.62	1.1%	16472	
2	0.707	F 0.83	16.8%	250318	Acetylated F peak
3	0.842	F 0.99	75.9%	1126815	Consistent with F
4	1.220	A 1.01	6.2%	91965	A peak
Total Area:				1485570	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range