


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well D07, 493015
[C968976DCC00EEEE] Jul 17, 2026 15:42:52 Pressure = 83 bar (83 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.687	F	0.82	19.7%	407145	Acetylated F peak
2	0.812	F	0.97	70.1%	1451606	Consistent with F
3	1.152	A	1.00	10.2%	211188	A peak
Total Area: 2069939						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range