


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2025, Rack A, Plate 1, Well G05, 171655
[8C729578B3020662] Dec 23, 2025 13:40:38 Pressure = 74 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	1.6%	12103	
2	0.493	F 0.57	1.0%	7742	
3	0.713	F 0.82	11.6%	87544	Acetylated F peak
4	0.845	F 0.97	53.5%	404839	Consistent with F
5	0.998	F 1.15	2.3%	17051	
6	1.223	A 1.01	30.1%	227795	A peak - REVIEW
Total Area:				757074	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range