


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well G04, 493087
[DD28950AE30A66C2] Jul 09, 2026 12:54:50

Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.683	F	0.82	16.0%	238588	Acetylated F peak
2	0.817	F	0.98	56.8%	844834	Consistent with F
3	1.157	A	1.00	27.1%	403124	A peak - REVIEW
Total Area: 1486546						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range