


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well D08, 493860
[C92DEC2784C53915] Jun 10, 2026 14:30:27 Pressure = 67 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.505	F 0.60	2.4%	18367	
2	0.697	F 0.83	15.8%	122539	Acetylated F peak
3	0.842	F 1.00	68.5%	532789	Consistent with F
4	1.193	A 0.99	13.3%	103655	A peak

Total Area: 777350

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range