


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2110, Rack A, Plate 1, Well H08, 498235
[892345415CFE244A] Mar 31, 2026 14:11:44 Pressure = 58 bar (57 to 59)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.418 | F 0.47 | 1.5% | 10947 | |
| 2 | 0.510 | F 0.57 | 1.1% | 8240 | |
| 3 | 0.728 | F 0.81 | 15.5% | 113720 | Acetylated F peak |
| 4 | 0.883 | F 0.99 | 60.7% | 444428 | Consistent with F |
| 5 | 1.245 | A 1.00 | 21.1% | 154595 | A peak - REVIEW |
| Total Area: | | | | 731930 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range