


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2098, Rack A, Plate 1, Well C07, 498855
[98762A5403E42448] Mar 19, 2026 12:16:33 Pressure = 60 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.417	F 0.47	1.3%	11264	
2	0.507	F 0.57	0.9%	8206	
3	0.718	F 0.81	12.0%	107013	Acetylated F peak
4	0.862	F 0.97	59.6%	532748	Consistent with F
5	1.250	A 1.01	26.2%	234478	A peak - REVIEW
Total Area:				893709	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range