


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2039, Rack A, Plate 1, Well A07, 171730-P  
 [D97789F72FB7D9B5] Jan 09, 2026 13:53:19 Pressure = 80 bar (77 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.21	4.9%	96327	
2	0.540	F 0.28	8.6%	170823	
3	0.887	F 0.46	13.0%	256292	
4	1.067	F 0.56	4.4%	86472	Acetylated F peak
5	1.867	F 0.98	38.3%	756338	Consistent with F
6	2.608	F 1.37	2.0%	40184	
7	2.745	F 1.44	2.2%	43485	
8	3.063	A 0.83	0.3%	5223	
9	3.633	A 0.99	25.5%	504218	A peak - REVIEW
10	4.068	A 1.10	0.2%	3453	
11	4.308	A 1.17	0.3%	6353	
12	4.483	S 0.86	0.4%	7650	
				Total Area: 1976818	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range