


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2176, Rack A, Plate 1, Well C08, 486004
[C978BD142B30CCE2] Jun 12, 2026 12:53:34 Pressure = 68 bar (68 to 70)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.423	F	0.50	0.7%	6104	
2	0.508	F	0.60	0.5%	4236	
3	0.697	F	0.83	9.3%	82028	Acetylated F peak
4	0.842	F	1.00	54.6%	481478	Consistent with F
5	1.185	A	0.99	35.0%	308377	A peak - REVIEW
Total Area:					882223	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range