


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2200, Rack A, Plate 1, Well A06, 487132  
[D83D826DCE08EEEE] Jul 17, 2026 14:26:04 Pressure = 82 bar (82 to 85)

FA



| PEAK        | RT    | REL RT | % CONC | AREA   | COMMENT           |
|-------------|-------|--------|--------|--------|-------------------|
| 1           | 0.515 | F 0.62 | 2.1%   | 15664  |                   |
| 2           | 0.693 | F 0.84 | 13.1%  | 95803  | Acetylated F peak |
| 3           | 0.823 | F 0.99 | 56.1%  | 411633 | Consistent with F |
| 4           | 1.150 | A 1.00 | 28.7%  | 210047 | A peak - REVIEW   |
| Total Area: |       |        |        | 733147 |                   |

Codes: 1) Wide A peak  
2) Area of A peak < 80%  
3) Peak area greater than expected  
4) Peak after A2  
5) A1c > 10%  
6) HbF or variant present  
7) Total sample area too small/big  
8) A2 is not within normal range