


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2022, Rack A, Plate 1, Well H07, 454131  
[C866C597887DF31D] Dec 19, 2025 13:13:53 Pressure = 73 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.47	1.0%	13823	
2	0.498	F 0.58	0.7%	10501	
3	0.725	F 0.85	16.3%	229113	Acetylated F peak
4	0.858	F 1.00	71.3%	1001353	Consistent with F
5	1.223	A 1.00	10.6%	148877	A peak
Total Area:				1403667	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range