


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2173, Rack A, Plate 1, Well E10, 503019
[CD6DF862FB18E468] Jun 10, 2026 14:58:45 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.507	F 0.60	2.6%	32151	
2	0.718	F 0.85	15.7%	197339	Acetylated F peak
3	0.833	F 0.99	67.9%	852916	Consistent with F
4	1.208	A 1.00	13.8%	173622	A peak

Total Area: 1256028

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range