


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2127, Rack A, Plate 1, Well C03, 173073
 [C9735DCAC8C9D395] Apr 21, 2026 10:37:15 Pressure = 64 bar (63 to 66)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|---------|-------------------|
| 1 | 0.512 | F 0.59 | 3.6% | 35699 | |
| 2 | 0.712 | F 0.83 | 20.9% | 210031 | |
| 3 | 0.852 | F 0.99 | 55.8% | 560031 | Consistent with F |
| 4 | 0.997 | F 1.16 | 4.6% | 46212 | |
| 5 | 1.237 | A 1.01 | 15.1% | 151721 | A peak |
| Total Area: | | | | 1003694 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range