


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2188, Rack A, Plate 1, Well D01, 501500  
[C83DC4C32ADD3B37] Jul 01, 2026 12:07:18 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.62	1.5%	15507	
2	0.695	F 0.83	13.6%	145184	Acetylated F peak
3	0.832	F 0.99	76.1%	812527	Consistent with F
4	1.163	A 1.01	8.9%	94772	A peak
Total Area:				1067990	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range