


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1998, Rack A, Plate 1, Well A09, 171211
[9967F967AECF5937] Nov 28, 2025 11:04:38 Pressure = 72 bar (71 to 73)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.385	F 0.45	2.6%	13881	
2	0.490	F 0.57	1.5%	7740	
3	0.695	F 0.81	9.7%	51689	Acetylated F peak
4	0.860	F 1.00	47.9%	255268	Consistent with F
5	1.015	F 1.18	2.6%	13636	
6	1.242	A 1.01	35.8%	190707	A peak - REVIEW
Total Area:				532921	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range