


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1359, Rack A, Plate 1, Well D03, 27479
[CDDCA04721851D35] Apr 03, 2024 10:51:51 Pressure = 61 bar (61 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.47	3.1%	24126	
2	0.480	F 0.54	5.4%	42656	
3	0.730	F 0.83	8.4%	66190	Acetylated F peak
4	0.880	F 1.00	62.0%	486616	Consistent with F
5	1.257	A 1.00	21.0%	164978	A peak - REVIEW
Total Area:				784566	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range